



COMMONWEALTH of VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CENTRAL LABORATORY
A Nationally Accredited Laboratory

700 NORTH 5TH STREET
RICHMOND, VIRGINIA 23219
(804) 786-6971

CERTIFICATE OF INSTRUMENT ACCURACY

INSTRUMENT: Intox EC/IR II
SERIAL NUMBER: 010442
CERTIFICATION DATE: February 26, 2014
DATE RECERTIFICATION DUE: August 25, 2014

The above listed instrument was found to meet all requirements for accuracy and performance established by the Department of Forensic Science as required by §9.1-1101 (B)(3). Measurement traceability is established through the use of reference materials supplied by an ISO/IEC 17025:2005 accredited calibration laboratory.

Estimation of Uncertainty of Measurement:

0.020 ± 0.004 g/210 L 0.150 ± 0.007 g/210 L
0.080 ± 0.004 g/210 L 0.250 ± 0.007 g/210 L

The Uncertainty of Measurement is reported at the 99.73% level of confidence and a coverage factor of k=3. The estimation of the Uncertainty of Measurement is calculated for the Certification process only.


Michelle M. Billeter
Forensic Scientist

CITY OF RICHMOND
COMMONWEALTH OF VIRGINIA

The foregoing was acknowledged before me this 4th day of March, 2014 by Michelle M. Billeter

Ciera N. Boxley, Notary Public

My commission expires: August 31, 2014



mmB
Jed

**Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 2/26/2014**

Instrument Serial Number:	010442	As of 26-Feb-14
----------------------------------	---------------	------------------------

Maintenance Date	Certification Date	Technician Initials	Remarks
26-Feb-14	26-Feb-14	MMB (26417)	CERTIFIED.
10-Dec-13	10-Dec-13	DBH (21250)	CALIBRATED BAROMETRIC PRESSURE SENSOR. REPLACED AND TESTED FUEL CELL MODULE. FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED.
15-May-13	15-May-13	DBH (21250)	REPLACED UPS BACK-UP BATTERY. CERTIFIED, REPLACED DRY GAS STANDARD.
29-Nov-12	29-Nov-12	SED (24178)	FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED. REPLACED DRY GAS STANDARD.
13-Jun-12	13-Jun-12	DBH (21250)	CERTIFIED, REPLACED DRY GAS STANDARD.
11-Jan-12	11-Jan-12	DBH (21250)	CERTIFIED, REPLACED DRY GAS STANDARD
21-Jul-11	21-Jul-11	DBH (21250)	CERTIFIED
26-Jan-11	26-Jan-11	RJP (22428)	REPLACED DRY GAS STANDARD. CERTIFIED.
10-Aug-10	10-Aug-10	RJP (22428)	CERTIFIED.
18-Feb-10	18-Feb-10	RJP (22428)	REPLACED DRY GAS STANDARD. CERTIFIED
25-Aug-09	25-Aug-09	RJP (22428)	CERTIFIED.
01-Apr-09	01-Apr-09	RJP (22428)	CERTIFIED.
25-Mar-09	25-Mar-09	NMJ (22392)	FIRMWARE/SOFTWARE UPGRADE, CERTIFIED.
20-Mar-09	20-Mar-09	NMJ (22392)	REPLACED TANK TOP LID, CERTIFIED.
06-Feb-09	16-Oct-08	RJP (22428)	REPLACED DRY GAS STANDARD.
16-Oct-08	16-Oct-08	RJP (22428)	CALIBRATED AND CERTIFIED.
30-Sep-08	30-Sep-08	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

mmB
nd

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010442 Worksheet Start Date 2/26/2014
Location DFS
Address
DFS Technician Michelle Billeter License No. 26417

☒ Laboratory ☐ On-Site

Site Specification: No detrimental environmental conditions exist. ☐

Instrument Barometer (mm HG) 756 Reference Barometer (mm HG) 756
Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/23/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.299
0.300	0.298	0.289	0.306	Sample 2	0.298
Precision		sample min	sample max	Sample 3	0.298
0.001		0.298	0.299		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.099
0.001		0.099	0.100		

Dry gas standard Lot No. (with tank no.) AG211702-09

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010442

Certification Date

2/26/2014

☐ Calibrated☒ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.251
0.250	0.249	0.241	0.256	Sample 2	0.249
Precision		sample min	sample max	Sample 3	0.248
0.003		0.248	0.251		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.148
0.150	0.149	0.144	0.153	Sample 2	0.148
Precision		sample min	sample max	Sample 3	0.148
0		0.148	0.148		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.079
0.080	0.080	0.077	0.083	Sample 2	0.079
Precision		sample min	sample max	Sample 3	0.079
0		0.079	0.079		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.019
0.020	0.020	0.017	0.023	Sample 2	0.019
Precision		sample min	sample max	Sample 3	0.019
0		0.019	0.019		

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bellet

Date

2/26/14

Issuing Analyst

Michelle Bellet

Date

2/27/14

MMB
Sed

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010442 Test Number: 926
Test Date: 02/26/2014 Test Time: 08:25 EST

Technician Initials: MMB

Dry Gas Target: 0.249

Lot Number: 0651225015-02 Exp Date: 04/01/2014

Tank Pressure: 286 psi Barometric Pressure: 756 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	08:25
BLK	0.000	08:26
CHK	0.251	08:27
BLK	0.000	08:28
CHK	0.249	08:29
BLK	0.000	08:30
CHK	0.248	08:31
BLK	0.000	08:32

Success

Standard Value (at sea level): 0.250 g/210L

Certification Status: Continued

nmb
jrd

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010442 Test Number: 226

Test Date: 02/26/2014 Test Time: 08:34 EST

Technician Initials: NMB

Dry Gas Target: 0.149

Lot Number: 0651215014-04 Exp Date: 04/01/2014

Tank Pressure: 659 psi Barometric Pressure: 756 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	08:35
BLK	0.000	08:35
CHK	0.148	08:36
BLK	0.000	08:37
CHK	0.148	08:38
BLK	0.000	08:39
CHK	0.148	08:40
BLK	0.000	08:41

Success

Standard Value (at sea level): 0.150 g/210L

Certification Status: Continued

mmB
Hd

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010442 Test Number: 127
Test Date: 02/26/2014 Test Time: 08:43 EST

Technician Initials: mmB

Dry Gas Target: 0.000

Lot Number: 0651208012-01 Exp Date: 04/01/2014

Tank Pressure: 845 psi Barometric Pressure: 756 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	08:43
BLK	0.000	08:44
CHK	0.079	08:44
BLK	0.000	08:46
CHK	0.079	08:46
BLK	0.000	08:47
CHK	0.079	08:48
BLK	0.000	08:49

Success

Standard Value (at sea level): 0.000 g/210L

Certification Status: Continued

mms

Ed

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010442 Test Number: 028
Test Date: 02/26/2014 Test Time: 08:51 EST

Technician Initials: MMS

Dry Gas Target: 0.020

Lot Number: 0651202011-04 Exp Date: 04/01/2014

Tank Pressure: 1012 psi Barometric Pressure: 756 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	08:52
BLK	0.000	08:53
CHK	0.013	08:54
BLK	0.000	08:55
CHK	0.013	08:56
BLK	0.000	08:57
CHK	0.013	08:57
BLK	0.000	08:59

Success

Standard Value (at sea level): 0.020 g/210L

Certification Status: Success

mmb
id

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010442 Test Number: 929

Test Date: 02/26/2014 Test Time: 09:04 EST

Dry Gas Target: 0.298

Lot Number: AG321804-04 Exp Date: 08/06/2015

Tank Pressure: 306 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:05
CHK	0.299	09:06
BLK	0.000	09:07
CHK	0.298	09:08
BLK	0.000	09:09
CHK	0.298	09:10

Test Status: *Success*

Calibration CRC: E671391D

mmB

FL

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010442 Test Number: 930

Test Date: 02/26/2014 Test Time: 09:17 EST

Dry Gas Target: 0.100

Lot Number: AG211702-09 Exp Date: 04/26/2014

Tank Pressure: 80 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:18
CHK	0.100	09:19
BLK	0.000	09:20
CHK	0.100	09:21
BLK	0.000	09:22
CHK	0.099	09:23

Test Status: *Success*

Calibration CRC: E671391D

mmB
her

INSTRUMENT, TEST,

DFS

BILLETER, MICHELLE, M

DFS Central Lab

26417

10/01/2014

02/26/2014

010442

02/26/2014

09:50 EST

0.00



IntoxNet MIS Report

Report Generated 26 Feb 2014 at 10:20

mmB
Jld

Test Results

Instrument Serial Number 010442

Test # 000931 Subject Test

Test Location 1 Department of
Test Date 26 Feb 2014

Test Location 2 Forensic Science
Test Time 09:43
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M
License Number 26417

Card Serial Number 126417

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 26 Feb 2014 End Time 09:51

Result Time 09:50

Result Date 26 Feb 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:44

Data Type BLK

Sample Value 0.000

Sample Time 09:44

Data Type CHK

Sample Value 0.100

Sample Time 09:45

Data Type BLK

Sample Value 0.000

Sample Time 09:46

Data Type SUBJ

Sample Value 0.000

Sample Time 09:47

Data Type BLK

Sample Value 0.000

Sample Time 09:48

Data Type SUBJ

Sample Value 0.000

Sample Time 09:50

Data Type BLK

Sample Value 0.000

Sample Time 09:51

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG211702-09

Standard Expiration Date 04/26/2014

Tank Pressure 75

Barometric Pressure 756 mmHg

Blow Sample Number 1 Blow Duration 3.71 sec

Blow Volume 1715 cc End-of-Blow Time 09:47

Blow Sample Number 2 Blow Duration 3.87 sec

Blow Volume 1696 cc End-of-Blow Time 09:50

Tamper Evident Stamp c35647a2

Test Status Code 0

Test Status Success